

**RIH – CHEST ANGIOGRAM
SIEMENS DEFINITION AS20 PROTOCOL**

Indications: Evaluation of the thoracic aorta

Position/Landmark	Head first or feet first-Supine 2cm superior to shoulders
Topogram Direction	Craniocaudal / Craniocaudal
Respiratory Phase	Inspiration
Scan Type	Helical
Ref kV/Ref mAs/Rotation time (sec) Pitch / Speed (mm/rotation) Safire Strength / Dose Optimization	Care kV 120 / Care Dose4D 150 / 0.5 sec 1.2:1 , 15.00mm 3 / 8
Detector width x Rows = Beam Collimation	0.625mm x 20 = 12.5mm
Average Tube Output	ctdi – 9 mGy dlp – 350 mGy.cm
Helical Set	body thickness/ recon part spacing algorithm recon destination .
Slice Thickness/ Spacing	1 chest cta 2mm x 2mm B30f medium smooth pacs
Algorithm	2 lungs 5mm x 5mm I70f very sharp pacs
Recon Destination	3 coronal chest 5mm x 5mm B30f medium smooth pacs
	4 thin chest .75mm x .7mm B30f medium smooth terarecon
Scan Start / End Locations	1cm superior to lung apices mid kidney
DFOV	38cm decrease appropriately
IV Contrast Volume / Type / Rate	100mL Iohexol (Omnipaque 350) / 4mL per second
Scan Delay	Bolus Tracking at the aortic arch
2D/3D Technique Used	Workstream 4D mpr of 5mm x 5mm coronal chest mip series, auto-transferred to PACS.
Comments: Recon 4 is a thin helical volume of the chest that is archived to the TeraRecon server.	
Images required in PACS	Topograms, 2mm x 2mm axial arterial chest, 5mm x 5mm coronal arterial chest, 5mm x 5mm axial lungs, Patient Protocol