

## RIH - SHOULDER CT GE LIGHTSPEED VCT PROTOCOL

**Indication: fracture, dislocation, osteomyelitis, bone injury, bone tumor.**

<b>Position/Landmark</b>	Supine , feet first Zero Appropriately																				
<b>Topogram Direction</b>	Craniocaudal																				
<b>Respiratory Phase</b>	Suspension																				
<b>Scan Type</b>	Helical																				
<b>KV / mA / Rotation time (sec) Pitch / Speed (mm/rotation) Noise Index / ASiR / Dose Reduction</b>	120kv / smart mA (100-450) / 0.5 sec 0.984:1 , 39.37mm 16.0 / 20 / 20%																				
<b>Detector width x Rows = Beam Collimation</b>	0.625mm x 64 = 40mm																				
<b>Helical Set</b>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">recon</th> <th style="text-align: left;">body part</th> <th style="text-align: left;">thickness/ spacing</th> <th style="text-align: left;">algorithm</th> <th style="text-align: left;">recon destination .</th> </tr> </thead> <tbody> <tr> <td>1</td> <td><b>shoulder bone</b></td> <td>2.5mm x 2.5mm</td> <td>bone</td> <td>pacs</td> </tr> <tr> <td>2</td> <td>thin shoulder</td> <td>.6mm x .6mm</td> <td>bone</td> <td>for dmpr</td> </tr> <tr> <td>3</td> <td><b>shoulder soft tissue</b></td> <td>2.5mm x 2.5mm</td> <td>standard</td> <td>pacs</td> </tr> </tbody> </table>	recon	body part	thickness/ spacing	algorithm	recon destination .	1	<b>shoulder bone</b>	2.5mm x 2.5mm	bone	pacs	2	thin shoulder	.6mm x .6mm	bone	for dmpr	3	<b>shoulder soft tissue</b>	2.5mm x 2.5mm	standard	pacs
recon	body part	thickness/ spacing	algorithm	recon destination .																	
1	<b>shoulder bone</b>	2.5mm x 2.5mm	bone	pacs																	
2	thin shoulder	.6mm x .6mm	bone	for dmpr																	
3	<b>shoulder soft tissue</b>	2.5mm x 2.5mm	standard	pacs																	
<b>Scan Start / End Locations</b>	determined by technologist or radiologist to include the anatomy of interest																				
<b>DFOV</b>	18cm decrease appropriately																				
<b>IV Contrast Volume / Type / Rate</b>	75mL Iohexol (Omnipaque 350) / 2mL per second if needed																				
<b>Scan Delay</b>	65 seconds																				
<b>2D/3D Technique Used</b>	DMPR of 3mm x 3mm <b>coronal and sagittal shoulder</b> series (auto-batch off), average mode, auto-transferred to PACS																				
<b>Comments:</b> Recon 1 is the 2.5mm x 2.5mm shoulder, bone algorithm ct going to PACS. Recon 2 is a single thin helical group of the shoulder for direct mpr. Recon 3 is the 2.5mm x 2.5mm shoulder, standard algorithm ct going to PACS.																					
<b>Images required in PACS</b>	Scouts, 2.5mm x 2.5mm axial shoulder bone, 2.5mm x 2.5mm axial shoulder standard, 3mm x 3mm sagittal shoulder, 3mm x 3mm coronal shoulder, Dose Report																				